



English

## Disposable Alligator Clip

### Intended Use

The Alligator Clip is intended for use in the recording of IONM [Intraoperative Neurophysiological Monitoring] as an accessory for general surgical instrumentation.

### Caution

Alligator Clips are for professional use only and should only be used in compliance with accepted industry standards. RhythmLink International, LLC is not responsible for injury, infection or other damage resulting from the use or misuse of this product. Product is sterile only if packaging is unopened.

### Instructions for Use

Remove the clip from the sterile pouch. Connect to monitoring equipment according to the equipment manufacturer's instructions.

This product is single-patient use only and should not be re-sterilized or reused. Discard clip after use.

- Reuse of this device may cause cross contamination and has NOT been validated for multiple uses.
- Re-sterilization of this device may cause unknown malfunctions. EO re-sterilization has not been validated for this device and may contribute to excessive EO gas residuals.



RhythmLink International, LLC  
1140 First Street South  
Columbia, SC, USA 29209-3540  
+1.866.633.3754 [toll-free]  
+1.803.252.1222  
+1.803.252.1111 [fax]  
sales@rhythmLink.com  
RhythmLink.com

STERILE EO Rx Only



To view a list of symbol definitions found on packaging and instructions for use, please visit [RhythmLink.com/symbols](http://RhythmLink.com/symbols).

RhythmLink® is a registered trademark of RhythmLink International, LLC.

L186001 rev006