



**RhythmLink International, LLC**  
1140 First Street South  
Columbia, SC 29209  
[RhythmLink.com](http://RhythmLink.com)

866.633.3754 (toll-free)  
803.252.1222  
803.252.1111 (fax)  
[sales@rhythmLink.com](mailto:sales@rhythmLink.com)

## Disposable Paired Sticky Pad™ Surface Electrodes

### Materials

- Pad - Conductive Vinyl layer coated with Cotton Micropore
- Pad Backing - Mylar
- Stimulating Hydrogel - AG702 Stimulating
- Recording Hydrogel - AG603 Sensing
- Ground Hydrogel - AG603 Sensing
- Leadwire - Tinned Copper with PVC Jacket
- Leadwire Pairing - Chemically Bonded
- Connector - DIN 42 802
- Sterilized - No

### Product Specifications

- Pad Sizes - 15 x 20 & 20 x 25 mm
- Leadwire Lengths - 1.0, 1.5 & 2.5 m

### Environment

- Latex-Free Electrode
- PVC-Free Packaging

### Packaging

- Pouch - Metallized Foil
- Pouch Size - 130 x 150 mm
- Boxes - Cardboard

### Additional Benefits

- Innovative leadwire glue enables custom separations as little or as much as needed
- Stimulating pads feature a specially formulated stimulating hydrogel that minimizes stimulation risks
- Recording pads feature a specially formulated recording hydrogel that optimizes signal recording
- Identifying letters distinguish between pads
- All Electrodes feature molded safety connectors with ample built-in strain relief

RhythmLink's Disposable Paired Sticky Pad Surface Electrodes are cleared for both stimulating and recording and are available in two varieties: Stimulating and Recording. These versatile, robust pads feature an aggressive hydrogel which keeps the Electrodes in place and enhances physiologic recordings or stimulations.

Visit [RhythmLink.com/stickypad](http://RhythmLink.com/stickypad) for additional information.

**Directions for Use** - Select appropriate Sticky Pad Electrode type and size. Remove excess hair, oils and dirt. Prep application site, remove Mylar backing from Electrode and place on application site. Apply pressure to center of Electrode and move to edges. Remove by pulling directly on Electrode. **DO NOT PULL ON THE CABLE!** Remove remaining hydrogel with clean, soapy water.

**Cleaning & Disinfecting Instructions** - This product is single-patient use only. Discard Electrode after use.

**Other Sticky Pad Varieties** - Single Stimulating, Single Recording and Ground Sticky Pads

